

Sensing Technology

WeatherStation® Instrument Calibration Certificate

Model:

PB200

ID.

1985270

Calibration Date:

5/25/2010

Manufacturer:

Airmar Technology Corporation, Milford NH USA

Test Procedure:

96-154-01

The WeatherStation® Instrument passes testing for IPX-6 heavy seas water standards and the IEC 60945 standards for exposed units₁. All units are calibrated in an onsite wind tunnel₂.

Unit Test Results:

Test Performed	Logged Value	Test Requirement
0 Knot Wind	0.7	Less than 1 Knot Max
50 Knot Cal ₄	N/A	Within ±2 Knots
Humidity	N/A	Within ±5 % RH
Temperature 5	22.3	Within ±1.1 °C
Pressure	1009.7	Within ±20 mBar
Rate Gyro – Dynamic 6	N/A	Under 4 deg/sec average error
Rate Gyro – Static	11.700000	Within ±40 Deg/min
Pitch	0.0	Within 1°
Roll	0.0	Within 1°

Ambient (Local) Test Conditions 3:

Humidity:

54.0 %RH

Temperature:

22.9 Deg C

Pressure:

1009.7 mBars

Equipment List

- Wind Tunnel Calibration
 - United Sensor Corp. Pitot Tube Calibrated to NIST traceable pitot tube
 - MKS Instruments Pressure Sensor
 - o BK Precision® 5491A Multimeter
- Ambient Temperature Readings (Temperature, humidity and pressure)
 - o Vaisala® PTU200
- Rate Gyro Testing
 - o SEI A2 Absolute encoder
 - US Digital R164 Motor Controller
- 1 After multiple heat cycles above 65° C (149°F) wind anemometer performance may require recalibration to remain within specifications at wind speeds above 50 knots (92.6 km/h).
- 2 The on site wind tunnel is calibrated with a pitot tube, which in turn was calibrated with a NIST traceable pitot tube.
- 3 Ambient conditions measured with a Vaisala PTU300. Temperature, humidity and pressure readings compared to Vaisala PTU300. The temperature, pressure and humidity readings from the Vaisala® instruments were calibrated by Vaisala® against instruments calibrated to NIST traceable instruments.
- 4 Units are calibrated at 50 knots with a full circle multipoint calibration.
- 5 Temperature recorded with 4+ knots of wind present.
- 6 Test conducted at 70 degrees per second.